

GP Temperature Version 1.2

Expression 5: extensive ($K \rightarrow K$)

Presentation of the different Good Practice (GP) expressions:

- | | |
|-----------------------|----------------------------|
| 1. simplified | ($K \rightarrow K$) |
| 2. typical | ($K \rightarrow K$) |
| 3. typical_adjustment | ($K \rightarrow K$) |
| 4. resistance | ($K \rightarrow \Omega$) |
| 5. extensive | ($K \rightarrow K$) |

The GP examples are formulated very generally and the numerical values given are arbitrary. The calibration laboratory must decide for itself whether all requirements for the representation of the calibration result are covered with these GP examples.

It cannot be deduced from the GP examples that they fulfil all requirements vis-à-vis third parties (e.g. to the needs of the customer). This must be agreed with the third party. In the same way, the creator of the DCC must ensure that the framework conditions given to him remain fulfilled.

Conformity statements are an essential part in factory calibrations. They generally do not occur in calibration certificates from national metrological state institutes. The DCC serves all representations of calibration results. Thus, the calibration laboratory must decide whether or not to include conformity statements. Therefore, conformity statements are optional in the DCC scheme.

Calibration certificates typically contain important text passages. Without these text passages, the certificate is of little or no significance. In this document, the text passages were deliberately omitted because the text passages are very individual and therefore cannot be reproduced here. The text passages reproduced in the XML files are also exemplary and must therefore be adapted by the user.

The page with the administrative data - with exception of the statements - is not dealt with here.

With the specifications 1, 2, 3 and 5, the temperature value on the calibration object is read during calibration. With specification 4 (resistance), the resistance of the calibration object is measured.

We would like to point out that the characteristics are intended to illustrate the variability of the GP temperature. Thus - depending on the calibration activity - the resistance can be measured on the calibration object or the temperature can be measured via an indicator. We see it as the task of the corresponding DKD technical committee to further develop this first draft for GP examples. The DCC team of PTB will be pleased to provide advice.

administrativeData

statements

statement

The results refer only to the object calibrated in this DCC. The measurement results are valid at the time of calibration. The applicant is responsible for arranging a recalibration in due time.

ISO/IEC 17025:2018-03 7.8.4.3

statement refType="basic_validityRange"

Lower limit: 306 K

Upper limit: 593 K

statement refType="basic_conformity"

The conformity statement is made on the basis of the customer's specifications. They are listed in the DCC.

Determination by "Customer GmbH"

PASS

statement refType="basic_recalibration"

Date when the calibration item is to be recalibrated at the latest after specification by the customer: 1959-10-22

Determination by "Customer GmbH"

measurementResults measurementResult

usedMethods

usedMethod refType="basic_uncertainty"

Expanded measurement uncertainty

GUM

usedMethod refType="gp_temperatureSensor"

Calibration of temperature sensors: ...

DKD-R 5-1:2018

usedMethod id="gp_uM1"

Additional method 1...

usedMethod id="gp_uM2"

Additional method 2...

usedMethod id="uM_totalSampleTime"

Recording time of individual measurements: ...

usedMethod id="uM_numberOfSamples"

Number of single measurements: ...

usedSoftware

software id="uS_1"

Description software 1: ... Release; R1.3

software id="uS_2"

Description software 2: ... Release; R4.0

measuringEquipments

measuringEquipment id="mE_1"

Description Pt 100 resistance thermometer 1: ...
Serial number, ...

measuringEquipment id="mE_2"

Description Pt 100 resistance thermometer 2: ...
Analogue calibration certificate
Serial number; Reference to calibration certificate of the measuringEquipment

measuringEquipment id="mE_3"

Description DAkKs - reference standard (PT25): ...
DCC
Serial number; Reference to calibration certificate including unique hash value of
measuringEquipment

influenceConditions

influenceCondition refType="gp_immersionDepth"

Immersion depth: ...
0.1 m

influenceCondition **refType="basic_temperature"**

Ambient condition temperature: ... 293 – 299 K

influenceCondition **refType="basic_humidityRelative"**

Ambient condition relative humidity: ... 0.2 – 0.7

influenceCondition **refType="basic_adjustment"**

<dcc:status>beforeAdjustment</dcc:status>

Adjustment

DRAFT

data refType="gp_table1"

| Start time | refType_area1 | basic_referenceValue | | | | | |
|-------------------|---------------|----------------------|--------------|---------------|---------------------|-------------------|------------------------|
| | | Method No. | Software No. | Equipment No. | Reference value / K | Measuring time | Calibration value / K |
| | refType_area2 | gp_uM* | gp_uS* | gp_mE* | | | basic_calibrationValue |
| | refType_area3 | | | | | | |
| 1957-08-13T08:00Z | | 1 | 1 | 1 | 306.245 | 1957-08-13T08:15Z | 306 |
| 1957-08-13T09:00Z | | 1 | 1 | 2 | 373.127 | 1957-08-13T09:15Z | 373 |
| 1957-08-13T10:00Z | | 1 | 2 | 2 | 448.249 | 1957-08-13T10:15Z | 448 |
| 1957-08-13T11:00Z | | 2 | 2 | 3 | 523.321 | 1957-08-13T11:15Z | 523 |
| 1957-08-13T12:00Z | | 2 | 2 | 3 | 593.151 | 1957-08-13T12:15Z | 593 |

Table continued ...

| refType_area1 | basic_measuredValue | | basic_measurementError | | | | | | |
|---------------|---------------------|-------------------|------------------------|---------------------------|---|------|----------------------------|------------------------|----|
| | Measured value/ K | Measuring time | Measurement error / K | U (Measurement error) / K | | | Total measurement time / s | Number of measurements | |
| refType_area2 | | | | | | | | | |
| refType_area3 | | | | | | | | | |
| | 306.35 | 1957-08-13T08:25Z | 0.105 | 0.89 | 2 | 0.95 | normal | 1262.38 | 60 |
| | 373.45 | 1957-08-13T09:25Z | 0.323 | 0.89 | 2 | 0.95 | normal | 1255.78 | 60 |
| | 448.05 | 1957-08-13T10:25Z | -0.199 | 0.89 | 2 | 0.95 | normal | 1258.11 | 60 |
| | 523.05 | 1957-08-13T11:25Z | -0.271 | 0.89 | 2 | 0.95 | normal | 1260.56 | 60 |
| | 593.35 | 1957-08-13T12:25Z | 0.199 | 0.89 | 2 | 0.95 | normal | 1264.45 | 60 |

Table continued ...

| refType_area1 | basic_conformity | basic_measurementError | | | |
|---------------|------------------|---------------------------|----------------------------|----------------------------|---------------------------|
| | | Lower tolerance limit / K | Lower acceptance limit / K | Upper acceptance limit / K | Upper tolerance limit / K |
| refType_area2 | | | | | |
| refType_area3 | | basic_toleranceLimitLower | basic_acceptanceLimitLower | basic_acceptanceLimitUpper | basic_toleranceLimitUpper |
| | pass | -1.55 | -0.75 | 0.75 | 1.55 |
| | pass | -1.55 | -0.75 | 0.75 | 1.55 |
| | pass | -1.55 | -0.75 | 0.75 | 1.55 |
| | pass | -2.05 | -0.90 | 0.90 | 2.05 |
| | pass | -2.05 | -0.90 | 0.90 | 2.05 |

results

result refType="gp_measuringResult1"

Measuring results

data refType="gp_table2"

| Start time | refType_area1 | basic_referenceValue | | | | | |
|-------------------|---------------|----------------------|--------------|---------------|---------------------|-------------------|------------------------|
| | | Method No. | Software No. | Equipment No. | Reference value / K | Measuring time | Calibration value / K |
| | refType_area2 | gp_uM* | gp_uS* | gp_mE* | | | basic_calibrationValue |
| | refType_area3 | | | | | | |
| 1957-08-13T13:00Z | | 1 | 1 | 1 | 306.248 | 1957-08-13T13:15Z | 306 |
| 1957-08-13T14:00Z | | 1 | 1 | 2 | 373.121 | 1957-08-13T14:15Z | 373 |
| 1957-08-13T15:00Z | | 1 | 2 | 2 | 448.253 | 1957-08-13T15:15Z | 448 |
| 1957-08-13T16:00Z | | 2 | 2 | 3 | 523.319 | 1957-08-13T16:15Z | 523 |
| 1957-08-13T17:00Z | | 2 | 2 | 3 | 593.154 | 1957-08-13T17:15Z | 593 |

Table continued ...

| refType_area1 | basic_measuredValue | | basic_measurementError | | | | | | |
|---------------|---------------------|-------------------|------------------------|---------------------------|---|------|----------------------------|------------------------|----|
| | Measured value/ K | Measuring time | Measurement error / K | U (Measurement error) / K | | | Total measurement time / s | Number of measurements | |
| refType_area2 | | | | | | | | | |
| refType_area3 | | | | | | | | | |
| | 306.32 | 1957-08-13T13:25Z | 0.072 | 0.061 | 2 | 0.95 | normal | 1261.23 | 60 |
| | 373.21 | 1957-08-13T14:25Z | 0.089 | 0.061 | 2 | 0.95 | normal | 1259.44 | 60 |
| | 448.36 | 1957-08-13T15:25Z | 0.107 | 0.061 | 2 | 0.95 | normal | 1257.27 | 60 |
| | 523.31 | 1957-08-13T16:25Z | -0.009 | 0.061 | 2 | 0.95 | normal | 1260.87 | 60 |
| | 593.07 | 1957-08-13T17:25Z | -0.084 | 0.061 | 2 | 0.95 | normal | 1264.59 | 60 |

Table continued ...

| refType_area1 | basic_conformity | basic_measurementError | | | |
|---------------|------------------|---------------------------|----------------------------|----------------------------|---------------------------|
| | | Lower tolerance limit / K | Lower acceptance limit / K | Upper acceptance limit / K | Upper tolerance limit / K |
| refType_area2 | | | | | |
| refType_area3 | | basic_toleranceLimitLower | basic_acceptanceLimitLower | basic_acceptanceLimitUpper | basic_toleranceLimitUpper |
| | pass | -0.58 | -0.23 | 0.23 | 0.58 |
| | pass | -0.58 | -0.23 | 0.23 | 0.58 |
| | pass | -0.58 | -0.23 | 0.23 | 0.58 |
| | pass | -0.78 | -0.30 | 0.30 | 0.78 |
| | pass | -0.78 | -0.30 | 0.30 | 0.78 |

measurementMetaData

metaData

declaration

Notes:

[1]: This is a typical procedure in the case of calibrating a temperature-sensor.